NiTi SuperFlex™ Tips (Patented)

- Instructions for use: 1) Rotate needle tip hub firmly on prefilled luer lock syringe
 - 2) Gently depress plunger to deliver solution
 - 3) Remove tip from syringe



To clean tip: Do not allow tips to exceed 125°C.

- 1. Soak in a mild, pH-neutral detergent until cleaning can be performed.
- 2. Rinse tips under a hard stream of water to remove debris on the outside surface of the tip.
- 3. Flush tip with a syringe filled with distilled water or alcohol to clear the tip.
- 4. Clean tip in an ultrasonic unit according to unit manufacturer's instructions using a low-foaming pH-neutral detergent. Use distilled water if possible. Do not allow tips to contact any other metals. Thoroughly rinse tips according to instructions supplied by detergent manufacturer.
- 5. Wrap tips for sterilization with wraps washed in pH-neutral detergents or disinfectants. Do not use commercial wraps with sterilization indicators that leave stains. Do not mix tips with other metals during autoclaving, or allow them to directly touch metal trays. Steam autoclave at 121°C/250°F, 15-20 psig for 30 minutes.



Emergo Europe Prinsessegracht 20 2514 AP The Hague The Netherlands 405107-I-ENG (3) Made in China

NiTi SuperFlex™ Tips (Patented)

Instructions for use: 1) Rotate needle tip hub firmly on prefilled luer lock syringe

2) Gently depress plunger to deliver solution

3) Remove tip from syringe

To clean tip: Do not allow tips to exceed 125°C.

- 1. Soak in a mild, pH-neutral detergent until cleaning can be performed.
- 2. Rinse tips under a hard stream of water to remove debris on the outside surface of the tip.
- 3. Flush tip with a syringe filled with distilled water or alcohol to clear the tip.
- 4. Clean tip in an ultrasonic unit according to unit manufacturer's instructions using a low-foaming pH-neutral detergent. Use distilled water if possible. Do not allow tips to contact any other metals. Thoroughly rinse tips according to instructions supplied by detergent manufacturer.
- 5. Wrap tips for sterilization with wraps washed in pH-neutral detergents or disinfectants. Do not use commercial wraps with sterilization indicators that leave stains. Do not mix tips with other metals during autoclaving, or allow them to directly touch metal trays. Steam autoclave at 121°C/250°F, 15-20 psig for 30 minutes.



Inter-Med, Inc 2200 South St Racine, WI 53404 USA Tel: +1-262-636-9755



405107-I-FNG (3) Made in China

Vista

NiTi SuperFlex™ Tips (Patented)

Instructions for use: 1) Rotate needle tip hub firmly on prefilled luer lock syringe

- 2) Gently depress plunger to deliver solution
- 3) Remove tip from syringe

To clean tip: Do not allow tips to exceed 125°C.

- 1. Soak in a mild, pH-neutral detergent until cleaning can be performed.
- 2. Rinse tips under a hard stream of water to remove debris on the outside surface of the tip.
- 3. Flush tip with a syringe filled with distilled water or alcohol to clear the tip.
- 4. Clean tip in an ultrasonic unit according to unit manufacturer's instructions using a low-foaming pH-neutral detergent. Use distilled water if possible. Do not allow tips to contact any other metals. Thoroughly rinse tips according to instructions supplied by detergent manufacturer.
- 5. Wrap tips for sterilization with wraps washed in pH-neutral detergents or disinfectants. Do not use commercial wraps with sterilization indicators that leave stains. Do not mix tips with other metals during autoclaving, or allow them to directly touch metal trays. Steam autoclave at 121°C/250°F, 15-20 psig for 30 minutes.





405107-I-ENG (3) Made in China

NiTi SuperFlex™ Tips (Patented)

- Instructions for use: 1) Rotate needle tip hub firmly on prefilled luer lock syringe
 - 2) Gently depress plunger to deliver solution
 - 3) Remove tip from syringe

Vista

To clean tip: Do not allow tips to exceed 125°C.

- 1. Soak in a mild, pH-neutral detergent until cleaning can be performed.
- 2. Rinse tips under a hard stream of water to remove debris on the outside surface of the tip.
- 3. Flush tip with a syringe filled with distilled water or alcohol to clear the tip.
- 4. Clean tip in an ultrasonic unit according to unit manufacturer's instructions using a low-foaming pH-neutral detergent. Use distilled water if possible. Do not allow tips to contact any other metals. Thoroughly rinse tips according to instructions supplied by detergent manufacturer.
- 5. Wrap tips for sterilization with wraps washed in pH-neutral detergents or disinfectants. Do not use commercial wraps with sterilization indicators that leave stains. Do not mix tips with other metals during autoclaving, or allow them to directly touch metal trays. Steam autoclave at 121°C/250°F, 15-20 psig for 30 minutes.

Inter-Med, Inc 2200 South St.

Racine, WI 53404 USA Tel: +1-262-636-9755 EC REP Emergo Europe

Made in China



NiTi SuperFlex™ Tips (Patented)

Instructions for use: 1) Rotate needle tip hub firmly on prefilled luer lock syringe

- 2) Gently depress plunger to deliver solution
- 3) Remove tip from syringe

Vista

To clean tip: Do not allow tips to exceed 125°C.

- 1. Soak in a mild, pH-neutral detergent until cleaning can be performed.
- 2. Rinse tips under a hard stream of water to remove debris on the outside surface of the tip.
- 3. Flush tip with a syringe filled with distilled water or alcohol to clear the tip.
- 4. Clean tip in an ultrasonic unit according to unit manufacturer's instructions using a low-foaming pH-neutral detergent. Use distilled water if possible. Do not allow tips to contact any other metals. Thoroughly rinse tips according to instructions supplied by detergent manufacturer.
- 5. Wrap tips for sterilization with wraps washed in pH-neutral detergents or disinfectants. Do not use commercial wraps with sterilization indicators that leave stains. Do not mix tips with other metals during autoclaving, or allow them to directly touch metal trays. Steam autoclave at 121°C/250°F, 15-20 psig for 30 minutes.



Inter-Med. Inc. 2200 South St. Racine, WI 53404 USA Tel: +1-262-636-9755



405107-I-ENG (3) Made in China

NiTi SuperFlex™ Tips (Patented)

- Instructions for use: 1) Rotate needle tip hub firmly on prefilled luer lock syringe
 - 2) Gently depress plunger to deliver solution
 - 3) Remove tip from syringe

To clean tip: Do not allow tips to exceed 125°C.

- 1. Soak in a mild, pH-neutral detergent until cleaning can be performed.
- 2. Rinse tips under a hard stream of water to remove debris on the outside surface of the tip
- 3. Flush tip with a syringe filled with distilled water or alcohol to clear the tip
- 4. Clean tip in an ultrasonic unit according to unit manufacturer's instructions using a low-foaming pH-neutral detergent. Use distilled water if possible. Do not allow tips to contact any other metals. Thoroughly rinse tips according to instructions supplied by detergent manufacturer
- 5. Wrap tips for sterilization with wraps washed in pH-neutral detergents or disinfectants. Do not use commercial wraps with sterilization indicators that leave stains. Do not mix tips with other metals during autoclaving, or allow them to directly touch metal trays. Steam autoclave at 121°C/250°F, 15-20 psig for 30 minutes.

Vista

Inter-Med, Inc. 2200 South St. Racine, WI 53404 USA Tel: +1-262-636-9755

EC REP Emergo Europe
Prinsessegracht 20
2514 AP The Hague
The Netherlands 405107-I-ENG (3)

NiTi SuperFlex™ Tips (Patented)

Instructions for use: 1) Rotate needle tip hub firmly on prefilled luer lock syringe

- 2) Gently depress plunger to deliver solution
- 3) Remove tip from syringe

To clean tip: Do not allow tips to exceed 125°C.

- 1. Soak in a mild, pH-neutral detergent until cleaning can be performed.
- 2. Rinse tips under a hard stream of water to remove debris on the outside surface of the tip
- 3. Flush tip with a syringe filled with distilled water or alcohol to clear the tip
- 4. Clean tip in an ultrasonic unit according to unit manufacturer's instructions using a low-foaming pH-neutral detergent. Use distilled water if possible. Do not allow tips to contact any other metals. Thoroughly rinse tips according to instructions supplied by detergent manufacturer.
- 5. Wrap tips for sterilization with wraps washed in pH-peutral detergents or disinfectants Do not use commercial wraps with sterilization indicators that leave stains. Do not mix tips with other metals during autoclaving, or allow them to directly touch metal trays. Steam autoclave at 121°C/250°F, 15-20 psig for 30 minutes.

Vista

Inter-Med, Inc.

2200 South St acine, WI 53404 USA Tel: +1-262-636-9755

Emergo Europe
Prinsessegracht 20

2514 AP The Hagu 405107-I-FNG (3)

Œ Made in China

Vista

NiTi SuperFlex™ Tips (Patented)

Instructions for use: 1) Rotate needle tip hub firmly on prefilled luer lock syringe

- 2) Gently depress plunger to deliver solution
- 3) Remove tip from syringe

To clean tip: Do not allow tips to exceed 125°C.

- 1. Soak in a mild, pH-neutral detergent until cleaning can be performed.
- 2. Rinse tips under a hard stream of water to remove debris on the outside surface of the tip
- 3. Flush tip with a syringe filled with distilled water or alcohol to clear the tip.
- 4. Clean tip in an ultrasonic unit according to unit manufacturer's instructions using a low-foaming pH-neutral detergent. Use distilled water if possible. Do not allow tips to contact any other metals. Thoroughly rinse tips according to instructions supplied by detergent manufacturer.
- 5. Wrap tips for sterilization with wraps washed in pH-neutral detergents or disinfectants. Do not use commercial wraps with sterilization indicators that leave stains. Do not mix tips with other metals during autoclaving, or allow them to directly touch metal travs. Steam autoclave at 121°C/250°F, 15-20 psig for 30 minutes.



EC REP Emergo Europe
Prinsessegracht 20 Prinsessegracht 20
2514 AP The Hague
The Netherlands



NiTi SuperFlex™ Tips (Patented)

- Instructions for use: 1) Rotate needle tip hub firmly on prefilled luer lock syringe
 - 2) Gently depress plunger to deliver solution
 - 3) Remove tip from syringe

Vista

To clean tip: Do not allow tips to exceed 125°C.

- 1. Soak in a mild, pH-neutral detergent until cleaning can be performed.
- 2. Rinse tips under a hard stream of water to remove debris on the outside surface of the tip
- 3. Flush tip with a syringe filled with distilled water or alcohol to clear the tip.
- 4. Clean tip in an ultrasonic unit according to unit manufacturer's instructions using a low-foaming pH-neutral detergent. Use distilled water if possible. Do not allow tips to contact any other metals. Thoroughly rinse tips according to instructions supplied by detergent manufacturer.
- 5. Wrap tips for sterilization with wraps washed in pH-neutral detergents or disinfectants. Steam autoclave at 121°C/250°F, 15-20 psig for 30 minutes.

Inter-Med, Inc. 2200 South St. acine, WI 53404 USA Tel: +1-262-636-9755 EC REP Emergo Europe Prinsessegracht 20

2514 AP The Hagu The Netherlands

405107-I-ENG (3)

Made in China

Œ

Do not use commercial wraps with sterilization indicators that leave stains. Do not mix tips with other metals during autoclaving, or allow them to directly touch metal trays.

NiTi SuperFlex™ Tips (Patented)

Instructions for use: 1) Rotate needle tip hub firmly on prefilled luer lock syringe

- 2) Gently depress plunger to deliver solution
- 3) Remove tip from syringe

To clean tip: Do not allow tips to exceed 125°C.

- 1. Soak in a mild, pH-neutral detergent until cleaning can be performed.
- 2. Rinse tips under a hard stream of water to remove debris on the outside surface of the tip.
- 3. Flush tip with a syringe filled with distilled water or alcohol to clear the tip.
- 4. Clean tip in an ultrasonic unit according to unit manufacturer's instructions using a low-foaming pH-neutral detergent. Use distilled water if possible. Do not allow tips to contact any other metals. Thoroughly rinse tips according to instructions supplied by detergent manufacturer.
- 5. Wrap tips for sterilization with wraps washed in pH-neutral detergents or disinfectants. Do not use commercial wraps with sterilization indicators that leave stains. Do not mix tips with other metals during autoclaving, or allow them to directly touch metal trays. Steam autoclave at 121°C/250°F, 15-20 psig for 30 minutes.



Inter-Med. Inc. 2200 South St. Racine, WI 53404 USA Tel: +1-262-636-9755

EC REP Emergo Europe
Prinsessegracht 20

2514 AP The Hague The Netherlands 405107-I-ENG (3) Made in China