



VISTA APEX

Porcelain Etch

REF 502500 - Standard Kit

Indications:

Vista Porcelain Etch kit is to be used any time resin bonding to porcelain is required. Porcelain Etch is used for the preparation of porcelain crowns, veneers or inlays prior to cementation. It is also used in preparing fractured porcelain and porcelain-to-metal crown bonding.

Instructions for Use:

1. Isolate area to be etched with rubber dam.
2. Remove porcelain syringe and Spira-Flo™ Brush Tips from their packaging.
3. Remove syringe cap and attach tip securely to syringe.
4. If preparing fractured porcelain for bonding, first smooth off the surface of the porcelain being etched with a medium grit bur, removing the glazed surface. Rinse away all particles with water and dry prepared porcelain with moisture/oil-free dry air.
5. Apply Porcelain Etch with brush tip to the prepared porcelain surface.
6. Let stand for 60 seconds, then rinse with water and dry surface (completely removing etchant).
7. Etched porcelain should appear frosty white. If not, repeat steps 5 & 6.
8. Apply Best-Etch™ to remove salts and debris (5 seconds), rinse and dry.
9. Remove Silane syringe and Spira-Flo™ Brush tip from packaging.
10. Remove syringe cap and attach tip securely to syringe.
11. Brush Silane onto etched porcelain surface and let dry completely. If needed, dry with moisture/oil-free air.
12. Place bonding resin according to manufacturer's instructions.

** STORE AT ROOM TEMPERATURE **



WARNING

May cause eye, respiratory and skin irritation. Harmful if swallowed. Wear protective eye wear and gloves.



Avoid unnecessary skin contact. May be disposed of in landfill or incinerator in accordance with local, state and federal regulations. **FIRST AID: IF ON SKIN:** Wash with plenty of soap and water.

IF IN EYES: Flush with large quantities of water for fifteen (15) minutes. Consult a physician.

IF SWALLOWED: Drink two glasses of water or milk. Seek medical attention if a substantial amount has been ingested. Do not induce vomiting.



Inter-Med, Inc.

2200 South St.
Racine, WI 53404 USA
Tel: +1-262-636-9755
www.vistaapex.com



Rx ONLY

Made in USA 502500-I-V-ENG (10)