



Vista-Blue™

REF 317007

VISTA | APEX



Harmful if swallowed. Avoid contact with skin and eyes

Instructions for Use:

Description:

Vista Blue is a water soluble solution which is used for the identification of cracks, calcified canals, root canal orifices and carious dentin.

Method:

- 1. Remove unit dose applicator and reservoir from packaging
2. Gently twist and push grey micro applicator brush into reservoir.
3. Remove micro applicator brush
4. Apply Vista-Blue onto the tooth surface to be checked.
5. Wait 5 – 10 seconds and rinse with water
6. Identify the crack in tooth.
7. Re-dip micro brush as needed and repeat steps 4 – 6.
8. Dispose of unit dose applicator and reservoir after every patient

Storage:

- 1. May stain clothing and skin.
2. Avoid contact with skin and eyes
3. In the event of eye contact, flush with copious amounts of water and seek immediate medical attention.
4. If contact with clothing occurs, stain may be removed with mild detergent and warm water.

Inter-Med, Inc. 2200 South St., Racine, WI 53404 USA Tel: +1-262-636-9755 Made in USA

Rx ONLY 317007-I-V-ENG (3)

Emergo Europe Prinsessegracht 20 2514 AP The Hague The Netherlands Tel: +31 (0)70 345 8570 Fax: +31 (0)70 346 7299



Vista-Blue™

REF 317007

VISTA | APEX



Harmful if swallowed. Avoid contact with skin and eyes

Instructions for Use:

Description:

Vista Blue is a water soluble solution which is used for the identification of cracks, calcified canals, root canal orifices and carious dentin.

Method:

- 1. Remove unit dose applicator and reservoir from packaging
2. Gently twist and push grey micro applicator brush into reservoir.
3. Remove micro applicator brush
4. Apply Vista-Blue onto the tooth surface to be checked.
5. Wait 5 – 10 seconds and rinse with water
6. Identify the crack in tooth.
7. Re-dip micro brush as needed and repeat steps 4 – 6.
8. Dispose of unit dose applicator and reservoir after every patient

Storage:

- 1. May stain clothing and skin.
2. Avoid contact with skin and eyes
3. In the event of eye contact, flush with copious amounts of water and seek immediate medical attention.
4. If contact with clothing occurs, stain may be removed with mild detergent and warm water.

Inter-Med, Inc. 2200 South St., Racine, WI 53404 USA Tel: +1-262-636-9755 Made in USA

Rx ONLY 317007-I-V-ENG (3)

Emergo Europe Prinsessegracht 20 2514 AP The Hague The Netherlands Tel: +31 (0)70 345 8570 Fax: +31 (0)70 346 7299



Vista-Blue™

REF 317007

VISTA | APEX



Harmful if swallowed. Avoid contact with skin and eyes

Instructions for Use:

Description:

Vista Blue is a water soluble solution which is used for the identification of cracks, calcified canals, root canal orifices and carious dentin.

Method:

- 1. Remove unit dose applicator and reservoir from packaging
2. Gently twist and push grey micro applicator brush into reservoir.
3. Remove micro applicator brush
4. Apply Vista-Blue onto the tooth surface to be checked.
5. Wait 5 – 10 seconds and rinse with water
6. Identify the crack in tooth.
7. Re-dip micro brush as needed and repeat steps 4 – 6.
8. Dispose of unit dose applicator and reservoir after every patient

Storage:

- 1. May stain clothing and skin.
2. Avoid contact with skin and eyes
3. In the event of eye contact, flush with copious amounts of water and seek immediate medical attention.
4. If contact with clothing occurs, stain may be removed with mild detergent and warm water.

Inter-Med, Inc. 2200 South St., Racine, WI 53404 USA Tel: +1-262-636-9755 Made in USA

Rx ONLY 317007-I-V-ENG (3)

Emergo Europe Prinsessegracht 20 2514 AP The Hague The Netherlands Tel: +31 (0)70 345 8570 Fax: +31 (0)70 346 7299



Vista-Blue™

REF 317007

VISTA | APEX



Harmful if swallowed. Avoid contact with skin and eyes

Instructions for Use:

Description:

Vista Blue is a water soluble solution which is used for the identification of cracks, calcified canals, root canal orifices and carious dentin.

Method:

- 1. Remove unit dose applicator and reservoir from packaging
2. Gently twist and push grey micro applicator brush into reservoir.
3. Remove micro applicator brush
4. Apply Vista-Blue onto the tooth surface to be checked.
5. Wait 5 – 10 seconds and rinse with water
6. Identify the crack in tooth.
7. Re-dip micro brush as needed and repeat steps 4 – 6.
8. Dispose of unit dose applicator and reservoir after every patient

Storage:

- 1. May stain clothing and skin.
2. Avoid contact with skin and eyes
3. In the event of eye contact, flush with copious amounts of water and seek immediate medical attention.
4. If contact with clothing occurs, stain may be removed with mild detergent and warm water.

Inter-Med, Inc. 2200 South St., Racine, WI 53404 USA Tel: +1-262-636-9755 Made in USA

Rx ONLY 317007-I-V-ENG (3)

Emergo Europe Prinsessegracht 20 2514 AP The Hague The Netherlands Tel: +31 (0)70 345 8570 Fax: +31 (0)70 346 7299